

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LKCP 475-329		
Serial Number:	22885		
Capacity:	2500	LBS	
Calibration Date:	7/26/2006		
Excitation:	10	VDC	
Calibration Factor:	2.3122	MV/V	Compression
Input Resistance:	387	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	60-160	Degrees F	

An output of 1.4645 MV/V is induced when a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	WHITE	(+)Signal
	GREEN	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 7/28/2006 Re-Calibration Date: 1 year after Shipment Date